



This is an amended version of report# 26-001677/D002.R000.

Reason: Updated product information.

Customer: The Hemp Collect
2014 SE 9th Ave
Portland Oregon 97214
United States of America (USA)

Product identity: Live D9 Caramel, Daytrip, Naturally Derived, 50mg

Material: Cannabinoid Edible

Laboratory ID: 26-001677-0002

Evidence of Cooling: No

Temp: 18.2 °C

Lot #: 5006SH_013026

Expiration Date (Client Provided): 02/17/2028

Serving Size #1: 18 g



**THE HEMP
COLLECT**

Sample Results

Potency		Method: J AOAC 2015 V98-6 (mod) ^b			Batch: 2601096		Analyzed: 02/10/26	
Analyte	Result	Units	LOQ	Notes	Serving Size #1			
					Result	Units	LOQ	
CBC	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
CBC-A	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
CBC-Total	< LOQ	%	0.0149		< LOQ	mg/18g	2.68	
CBD [±]	0.00864	%	0.0079		1.55	mg/18g	1.43	
CBD-A [±]	0.0108	%	0.0079		1.94	mg/18g	1.43	
CBD-Total [±]	0.0181	%	0.0149		3.26	mg/18g	2.68	
CBDV	0.0160	%	0.0079		2.88	mg/18g	1.43	
CBDV-A	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
CBDV-Total	0.0160	%	0.0148		2.88	mg/18g	2.66	
CBE	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
CBG	0.0276	%	0.0079		4.96	mg/18g	1.43	
CBG-A	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
CBG-Total	0.0276	%	0.0148		4.97	mg/18g	2.66	
CBL	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
CBL-A	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
CBL-Total	< LOQ	%	0.0149		< LOQ	mg/18g	2.68	
CBN	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
CBT	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
Δ10-THC-9R	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
Δ10-THC-9S	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
Δ10-THC-Total	< LOQ	%	0.0158		< LOQ	mg/18g	2.85	
Δ8-THC [±]	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
Δ8-THCV	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
Δ9-THC [±]	0.214	%	0.0079		38.4	mg/18g	1.43	
Δ9-THC-A [±]	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
Δ9-THC-Total [±]	0.214	%	0.0149		38.5	mg/18g	2.68	
Δ9-THCP	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	
Δ9-THCV	< LOQ	%	0.0079		< LOQ	mg/18g	1.43	


Potency Method: J AOAC 2015 V98-6 (mod)^b Batch: 2601096 Analyzed: 02/10/26

Analyte	Result	Units	LOQ	Notes	Serving Size #1		
					Result	Units	LOQ
Δ9-THCV-A	< LOQ	%	0.0079		< LOQ	mg/18g	1.43
Δ9-THCV-Total	< LOQ	%	0.0148		< LOQ	mg/18g	2.66
exo-THC	< LOQ	%	0.0079		< LOQ	mg/18g	1.43
Total Cannabinoids	0.277	%			49.9	mg/18g	

Microbiology

Analyte	Result	Limits	Units	LOQ	Batch	Analyzed Method	Status	Notes
Salmonella spp. [±]	Negative		/25g		2601048	02/11/26 AOAC 2020.02 ^b		
EHEC including STEC [±]	Negative		/25g		2601049	02/11/26 AOAC 2020.06 ^b		

Solvents Method: Residual Solvents by HS-GC-MS^b Units µg/g Batch 2601094 Analyze: 02/11/26

Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
1,4-Dioxane [±]	< LOQ	380	100	pass		2-Butanol [±]	< LOQ	5000	200	pass	
2-Ethoxyethanol [±]	< LOQ	160	30.0	pass		2-Methylbutane (Isopentane) [±]	< LOQ		200		
2-Methylpentane [±]	< LOQ		30.0			2-Propanol (IPA) [±]	< LOQ	5000	200	pass	
2,2-Dimethylbutane [±]	< LOQ		30.0			2,2-Dimethylpropane (neo-pentane) [±]	< LOQ		200		
2,3-Dimethylbutane [±]	< LOQ		30.0			3-Methylpentane [±]	< LOQ		30.0		
Acetone [±]	< LOQ	5000	200	pass		Acetonitrile [±]	< LOQ	410	100	pass	
Benzene [±]	< LOQ	2.00	1.00	pass		Butanes (sum) [±]	< LOQ	5000	400	pass	
Cyclohexane [±]	< LOQ	3880	200	pass		Ethyl acetate [±]	< LOQ	5000	200	pass	
Ethyl benzene	< LOQ		200			Ethyl ether [±]	< LOQ	5000	200	pass	
Ethylene glycol [±]	< LOQ	620	200	pass		Ethylene oxide [±]	< LOQ	50.0	20.0	pass	
Hexanes (sum) [±]	< LOQ	290	150	pass		Isopropyl acetate [±]	< LOQ	5000	200	pass	
Isopropylbenzene (Cumene) [±]	< LOQ	70.0	30.0	pass		m,p-Xylene [±]	< LOQ		200		
Methanol [±]	< LOQ	3000	200	pass		Methylene chloride [±]	< LOQ	600	60.0	pass	
Methylpropane (Isobutane) [±]	< LOQ		200			n-Butane [±]	< LOQ		200		
n-Heptane [±]	< LOQ	5000	200	pass		n-Hexane [±]	< LOQ		30.0		
n-Pentane [±]	< LOQ		200			o-Xylene [±]	< LOQ		200		
Pentanes (sum) [±]	< LOQ	5000	600	pass		Propane [±]	< LOQ	5000	200	pass	
Tetrahydrofuran [±]	< LOQ	720	100	pass		Toluene [±]	< LOQ	890	100	pass	
Total Xylenes [±]	< LOQ		400			Total Xylenes and Ethyl benzene	< LOQ	2170	600	pass	

Pesticides Method: AOAC 2007.01 & EN 15662 (mod) Units mg/kg Batch 2601204 Analyze: 02/16/26

Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
Abamectin [±]	< LOQ	0.50	0.250	pass		Acephate [±]	< LOQ	0.40	0.200	pass	
Acequinocyl [±]	< LOQ	2.0	1.00	pass		Acetamiprid [±]	< LOQ	0.20	0.100	pass	
Aldicarb [±]	< LOQ	0.40	0.200	pass		Azoxystrobin [±]	< LOQ	0.20	0.100	pass	
Bifenazate [±]	< LOQ	0.20	0.100	pass		Bifenthrin [±]	< LOQ	0.20	0.100	pass	
Boscalid [±]	< LOQ	0.40	0.200	pass		Carbaryl [±]	< LOQ	0.20	0.100	pass	
Carbofuran [±]	< LOQ	0.20	0.100	pass		Chlorantraniliprole [±]	< LOQ	0.20	0.100	pass	
Chlorfenapyr [±]	< LOQ	1.0	0.500	pass		Chlorpyrifos-ethyl [±]	< LOQ	0.20	0.100	pass	
Clofentezine [±]	< LOQ	0.20	0.100	pass		Cyfluthrin (sum) [±]	< LOQ	1.0	0.500	pass	
Cypermethrin (sum) [±]	< LOQ	1.0	0.500	pass		Daminozide [±]	< LOQ	1.0	0.500	pass	
Diazinon [±]	< LOQ	0.20	0.100	pass		Dichlorvos [±]	< LOQ	1.0	0.500	pass	



Pesticides		Method: AOAC 2007.01 & EN 15662 (mod)				Units mg/kg	Batch 2601204	Analyze: 02/16/26			
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
Dimethoate [±]	< LOQ	0.20	0.100	pass		Ethoprophos [±]	< LOQ	0.20	0.100	pass	
Etofenprox [±]	< LOQ	0.40	0.200	pass		Etoazole [±]	< LOQ	0.20	0.100	pass	
Fenoxycarb [±]	< LOQ	0.20	0.100	pass		Fenpyroximate [±]	< LOQ	0.40	0.200	pass	
Fipronil [±]	< LOQ	0.40	0.200	pass		Flonicamid [±]	< LOQ	1.0	0.400	pass	
Fludioxonil [±]	< LOQ	0.40	0.200	pass		Hexythiazox [±]	< LOQ	1.0	0.400	pass	
Imazalil [±]	< LOQ	0.20	0.100	pass		Imidacloprid [±]	< LOQ	0.40	0.200	pass	
Kresoxim-methyl [±]	< LOQ	0.40	0.200	pass		Malathion [±]	< LOQ	0.20	0.100	pass	
Metalaxyl [±]	< LOQ	0.20	0.100	pass		Methiocarb [±]	< LOQ	0.20	0.100	pass	
Methomyl [±]	< LOQ	0.40	0.200	pass		MGK-264 [±]	< LOQ	0.20	0.100	pass	
Myclobutanil [±]	< LOQ	0.20	0.100	pass		Naled [±]	< LOQ	0.50	0.250	pass	
Oxamyl [±]	< LOQ	1.0	0.500	pass		Paclobutrazole [±]	< LOQ	0.40	0.200	pass	
Parathion-methyl [±]	< LOQ	0.20	0.100	pass		Permethrin [±]	< LOQ	0.20	0.100	pass	
Phosmet [±]	< LOQ	0.20	0.100	pass		Piperonyl butoxide [±]	< LOQ	2.0	1.00	pass	
Prallethrin [±]	< LOQ	0.20	0.100	pass		Propiconazole [±]	< LOQ	0.40	0.200	pass	
Propoxur [±]	< LOQ	0.20	0.100	pass		Pyrethrin I (total) [±]	< LOQ	1.0	0.500	pass	
Pyridaben [±]	< LOQ	0.20	0.100	pass		Spinosad [±]	< LOQ	0.20	0.100	pass	
Spiromesifen [±]	< LOQ	0.20	0.100	pass		Spirotetramat [±]	< LOQ	0.20	0.100	pass	
Spiroxamine [±]	< LOQ	0.40	0.200	pass		Tebuconazole [±]	< LOQ	0.40	0.200	pass	
Thiacloprid [±]	< LOQ	0.20	0.100	pass		Thiamethoxam [±]	< LOQ	0.20	0.100	pass	
Trifloxystrobin [±]	< LOQ	0.20	0.100	pass							

Metals

Analyte	Result	Limits	Units	LOQ	Batch	Analyzed Method	Status	Notes
Arsenic [±]	< LOQ	0.200	mg/kg	0.0191	2601073	02/10/26 AOAC 2013.06 (mod.) ^p	pass	
Cadmium [±]	< LOQ	0.200	mg/kg	0.0191	2601073	02/10/26 AOAC 2013.06 (mod.) ^p	pass	
Lead [±]	< LOQ	0.500	mg/kg	0.0191	2601073	02/10/26 AOAC 2013.06 (mod.) ^p	pass	
Mercury [±]	< LOQ	0.100	mg/kg	0.00957	2601073	02/10/26 AOAC 2013.06 (mod.) ^p	pass	

Mycotoxins

Analyte	Result	Limits	Units	LOQ	Batch	Analyzed Method	Status	Notes
Aflatoxin B1 [±]	< LOQ		µg/kg	5.00	2601141	02/13/26 Mycotoxins by AOAC 2007.01		
Aflatoxin B2 [±]	< LOQ		µg/kg	5.00	2601141	02/13/26 Mycotoxins by AOAC 2007.01		
Aflatoxin G1 [±]	< LOQ		µg/kg	5.00	2601141	02/13/26 Mycotoxins by AOAC 2007.01		
Aflatoxin G2 [±]	< LOQ		µg/kg	5.00	2601141	02/13/26 Mycotoxins by AOAC 2007.01		
Ochratoxin A [±]	< LOQ	20.0	µg/kg	5.00	2601141	02/13/26 Mycotoxins by AOAC 2007.01	pass	
Total Aflatoxins	< LOQ	20.0	µg/kg	20.0		02/16/26 Mycotoxins by AOAC 2007.01 ^p	pass	

Notes:

Measurement of Uncertainty for D9 THC in edibles is +/- 7.5%.



Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

Threshold Note: OAR 333-007-0400

Ⓟ = ISO/IEC 17025:2017 accredited method.

⊕ = TNI accredited analyte.

Units of Measure

% wt = $\mu\text{g/g}$ divided by 10,000

/25g = Per 25g

$\mu\text{g/g}$ = Microgram per gram

$\mu\text{g/kg}$ = Micrograms per kilogram = parts per billion (ppb)

mg/kg = Milligram per kilogram = parts per million (ppm)

% = Percentage of sample

mg/18g = Milligram per 18g

Approved Signatory

Derrick Tanner
General Manager





12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 26-001677/D002.R001
Report Date: 04/13/2026
ORELAP#: OR100028
Received: 02/09/26 09:25



**Hemp & Cannabis
Chain of Custody**

**The-Hemp-
Collect-1770083080**

Company Details Company: <u>The Hemp Collect</u> Contact: <u>Sierra Solnick</u> Street Address: <u>2014 SE 9th Ave</u> City, State, Zip: <u>Portland, OR 97214</u> Email: <u>sierra@thehempcollect.com</u> Contact Phone: <u>8607520027</u> Billing Information Billing Email: <u>accounting@thehempcollect.com</u>				Project Details Turnaround Time: <u>5 Business Days Req. For Micro Testing Standard</u> Relinquishment Sampling, Courier & Shipping Options: <u>By Shipping Service (USPS, UPS, Fedex)</u> Receipt Information Evidence of Cooling?: <u>No</u> Sample Condition: <u>Satisfactory</u> Prelog Storage: <u>Canna Shelves</u>			Testing CH005 - Oregon Package
#	Sample Name	Lot Additional Sample ID	Material	Amount Provided	Reporting Unit	Specifications	
1	Live D9 Caramel, Anytime, Naturally Derived 42mg	5004SH_013026	Cannabinoid Edible	10 each	mg/g	Please report mg per 18 grams	✓
2	Live D9 Caramel, Daytrip, Naturally Derived, 50mg	5006SH_013026	Cannabinoid Edible	10 each	mg/g	Please report mg per 18 grams	✓

Package Details

Oregon Package: Aflatoxins+Ochratoxin | OLCC • Cannabis Heavy Metals Profile OR • Micro Profile OR (OLCC Comp) • Pesticides (OR - Cannabis) • Potency Cannabis (Basic+Expanded) • Residual Solvents (Cannabis - Oregon)

Relinquished By	Date	Time	Received By	Date	Time	Received Temp., °C	IR Therm. CL#
<i>Sierra Solnick</i>	<i>02/02/2026</i>	<i>17:44</i>	<i>dst</i>	<i>02/09/2026</i>	<i>09:25</i>	<i>18.20</i>	<i>CL-0530</i>

Samples submitted to Columbia Laboratories with testing requirements constitute an agreement for services in accordance with the [current terms of services](#) associated with this COC. By signing "Relinquished by" you are agreeing to these terms.

Columbia Laboratories
12423 NE Whitaker Way
Portland, OR 97230

P: (503) 254-1794
info.cj@tentamus.com

Page 1 of 1
www.columbialaboratories.com


 Revision: 2 Document ID: 7087
 Legacy ID: CFL-E33Effective:

Laboratory Quality Control Results

Residual Solvents				Batch ID: 2601094					
Method Blank				Laboratory Control Sample					
Analyte	Result	LOQ	Notes	Result	Spike	Units	% Rec	Limits	Notes
1,1-Dichloroethane	ND	< 1		1.02	1	µg/g	102.0	50-150	
1,1-Dichloroethene	ND	< 1		1.06	1	µg/g	106.0	50-150	
1,2-Dichloroethene, trans-	ND	< 1		1.09	1	µg/g	109.0	50-150	
1,4-Dioxane	ND	< 100		550	504	µg/g	109.1	60-120	
2,2-Dimethylbutane	ND	< 30		158	171	µg/g	92.4	60-120	
2,2-Dimethylpropane	ND	< 200		1120	956	µg/g	117.2	60-120	
2,3-Dimethylbutane	ND	< 30		120	166	µg/g	72.3	60-120	
2-Butanol	ND	< 200		1250	1620	µg/g	77.2	60-120	
2-Ethoxyethanol	ND	< 30		118	165	µg/g	71.5	60-120	
2-Methylbutane	ND	< 200		1340	1610	µg/g	83.2	60-120	
2-Methylpentane	ND	< 30		174	167	µg/g	104.2	60-120	
2-Propanol	ND	< 200		1340	1610	µg/g	83.2	60-120	
3-Methylpentane	ND	< 30		141	163	µg/g	86.5	60-120	
Acetone	ND	< 200		1370	1610	µg/g	85.1	60-120	
Acetonitrile	ND	< 100		404	506	µg/g	79.8	60-120	
Anisole	ND	< 500		1780	1620	µg/g	109.9	50-150	
Benzene	ND	< 1		1.02	1	µg/g	102.0	50-150	
Butane	ND	< 200		806	769	µg/g	104.8	60-120	
Chloroform	ND	< 1		1.06	1	µg/g	106.0	50-150	
Cumene	ND	< 30		180	162	µg/g	111.1	60-120	
Cyclohexane	ND	< 200		1670	1610	µg/g	103.7	60-120	
Dichloromethane	ND	< 1		1.12	1	µg/g	112.0	50-150	
DMSO	ND	< 500		1580	1620	µg/g	97.5	50-150	
Ethanol	ND	< 200		1390	1620	µg/g	85.8	60-120	
Ethyl acetate	ND	< 200		1310	1620	µg/g	80.9	60-120	
Ethyl Ether	ND	< 200		1480	1610	µg/g	91.9	60-120	
Ethylbenzene	ND	< 200		1070	969	µg/g	110.4	60-120	
Ethylene Glycol	ND	< 200		346	503	µg/g	68.8	60-120	
Ethylene Oxide	ND	< 1		0.861	1	µg/g	86.1	50-150	
Heptane	ND	< 200		1330	1610	µg/g	82.6	60-120	
Hexane	ND	< 30		158	166	µg/g	95.2	60-120	
Isobutane	ND	< 200		796	770	µg/g	103.4	60-120	
Isobutyl Acetate	ND	< 500		1580	1640	µg/g	96.3	50-150	
Isopropyl Acetate	ND	< 200		1360	1610	µg/g	84.5	60-120	
m,p-Xylene	ND	< 200		1090	994	µg/g	109.7	60-120	
Methanol	ND	< 200		1090	1620	µg/g	67.3	60-120	
Methylisobutylketone	ND	< 500		1550	1620	µg/g	95.7	50-150	
MTBE	ND	< 500		1800	1610	µg/g	111.8	50-150	
N,N-dimethylacetamide	ND	< 150		469	486	µg/g	96.5	50-150	
o-Xylene	ND	< 200		1100	981	µg/g	112.1	60-120	
Pentane	ND	< 200		1280	1610	µg/g	79.5	60-120	
Propane	ND	< 200		623	585	µg/g	106.5	60-120	
Propyl Acetate	ND	< 500		1590	1610	µg/g	98.8	50-150	
Sulfolane	ND	< 50		158	193	µg/g	81.9	50-150	
Tetrahydrofuran	ND	< 100		474	488	µg/g	97.1	60-120	
Toluene	ND	< 100		548	505	µg/g	108.5	60-120	
Triethylamine	ND	< 500		1590	1610	µg/g	98.8	50-150	


 Revision: 2 Document ID: 7087
 Legacy ID: CFL-E33Effective:

QC - Sample Duplicate
Sample ID: 26-001639-0001

Analyte	SR Result	SD Result	LOQ	Units	RPD	Limits	Accept/Fail	Notes
1,1-Dichloroethane	ND	ND	1	µg/g	0.0	< 20	Acceptable	
1,1-Dichloroethene	ND	ND	1	µg/g	0.0	< 20	Acceptable	
1,2-Dichloroethene, trans-	ND	ND	1	µg/g	0.0	< 20	Acceptable	
1,4-Dioxane	ND	ND	100	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylpropane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,3-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Butanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Ethoxyethanol	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Methylbutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Propanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
3-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Acetone	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Acetonitrile	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Anisole	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Benzene	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Butane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Chloroform	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Cumene	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Cyclohexane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Dichloromethane	ND	ND	1	µg/g	0.0	< 20	Acceptable	
DMSO	533	600	500	µg/g	11.8	< 20	Acceptable	
Ethanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Ether	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylbenzene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylene Glycol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylene Oxide	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Heptane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Hexane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Isobutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Isopropyl Acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
m,p-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Methanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Methylisobutylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
MTBE	ND	ND	500	µg/g	0.0	< 20	Acceptable	
N,N-dimethylacetamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
o-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Pentane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Propane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Propyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Sulfolane	ND	ND	50	µg/g	0.0	< 20	Acceptable	
Tetrahydrofuran	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Toluene	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Triethylamine	ND	ND	500	µg/g	0.0	< 20	Acceptable	

Abbreviations

 ND - None Detected at or above MRL
 RPD - Relative Percent Difference
 LOQ - Limit of Quantitation

Units of Measure:

µg/g - Microgram per gram or ppm



Revision: 4 Document ID: 7148
Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

J AOAC 2015 V98-6 **Batch ID: 2601096**

Laboratory Control Sample

Analyte	LCS	Result	Spike	Units	% Rec	Limits	Evaluation	Notes
CBDVA	2	0.0299	0.0303	%	98.9	80.0 - 120	Acceptable	
CBDV	2	0.0291	0.0293	%	99.1	80.0 - 120	Acceptable	
CBE	2	0.0325	0.0322	%	101	80.0 - 120	Acceptable	
CBDA	1	0.0251	0.0262	%	95.7	90.0 - 110	Acceptable	
CBGA	1	0.0271	0.0272	%	99.7	80.0 - 120	Acceptable	
CBG	1	0.0255	0.0257	%	98.9	80.0 - 120	Acceptable	
CBD	1	0.0246	0.0240	%	102	90.0 - 110	Acceptable	
THCV	2	0.0297	0.0298	%	99.7	80.0 - 120	Acceptable	
d8THCV	2	0.0307	0.0311	%	98.5	80.0 - 120	Acceptable	
THCVA	2	0.0307	0.0309	%	99.3	80.0 - 120	Acceptable	
CBN	1	0.0255	0.0252	%	101	80.0 - 120	Acceptable	
exo-THC	2	0.0277	0.0280	%	98.9	80.0 - 120	Acceptable	
d9THC	1	0.0251	0.0249	%	101	90.0 - 110	Acceptable	
d8THC	1	0.0261	0.0266	%	97.9	90.0 - 110	Acceptable	
9S-d10THC	1	0.0279	0.0279	%	99.8	80.0 - 120	Acceptable	
CBL	2	0.0298	0.0308	%	96.6	80.0 - 120	Acceptable	
9R-d10THC	1	0.0297	0.0300	%	99.3	80.0 - 120	Acceptable	
CBC	2	0.0302	0.0302	%	100	80.0 - 120	Acceptable	
THCA	1	0.0258	0.0262	%	98.2	90.0 - 110	Acceptable	
CBCA	2	0.0306	0.0316	%	96.9	80.0 - 120	Acceptable	
CBLA	2	0.0307	0.0314	%	97.9	80.0 - 120	Acceptable	
d9THCP	2	0.0319	0.0320	%	99.6	80.0 - 120	Acceptable	
CBT	2	0.0276	0.0278	%	99.2	80.0 - 120	Acceptable	

Method Blank

Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDVA	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBDV	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBE	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBDA	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBGA	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBG	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBD	<LOQ	0.00749	%	< 0.00749	Acceptable	
THCV	<LOQ	0.00749	%	< 0.00749	Acceptable	
d8THCV	<LOQ	0.00749	%	< 0.00749	Acceptable	
THCVA	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBN	<LOQ	0.00749	%	< 0.00749	Acceptable	
exo-THC	<LOQ	0.00749	%	< 0.00749	Acceptable	
d9THC	<LOQ	0.00749	%	< 0.00749	Acceptable	
d8THC	<LOQ	0.00749	%	< 0.00749	Acceptable	
9S-d10THC	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBL	<LOQ	0.00749	%	< 0.00749	Acceptable	
9R-d10THC	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBC	<LOQ	0.00749	%	< 0.00749	Acceptable	
THCA	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBCA	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBLA	<LOQ	0.00749	%	< 0.00749	Acceptable	
d9THCP	<LOQ	0.00749	%	< 0.00749	Acceptable	
CBT	<LOQ	0.00749	%	< 0.00749	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

% - Percent



Revision: 4 Document ID: 7148
Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

AOAC 2015 V98-6		Batch ID: 2601096						
Sample Duplicate		Sample ID: 26-001625-0001						
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDVA	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
CBDV	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
CBE	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
CBDA	<LOQ	<LOQ	0.00792	%	NA	< 10	Acceptable	
CBGA	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
CBG	0.0138	0.0141	0.00792	%	2.09	< 20	Acceptable	
CBD	0.505	0.507	0.00792	%	0.235	< 10	Acceptable	
THCV	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
d8THCV	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
THCVA	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
CBN	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
exo-THC	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
d9THC	0.0204	0.0206	0.00792	%	0.957	< 10	Acceptable	
d8THC	<LOQ	<LOQ	0.00792	%	NA	< 10	Acceptable	
9S-d10THC	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
CBL	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
9R-d10THC	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
CBC	0.0213	0.0213	0.00792	%	0.329	< 20	Acceptable	
THCA	<LOQ	<LOQ	0.00792	%	NA	< 10	Acceptable	
CBCA	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
CBLA	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
d9THCP	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	
CBT	<LOQ	<LOQ	0.00792	%	NA	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

% - Percent